

Rapid Test Detecting Betalactams, Sulfonamides And Tetracyclins In Milk




TRISENSOR is a rapid assay in dipstick format detecting the contamination of milk samples by Betalactams, Tetracyclines and Sulfonamides molecules.

The test can be used in-situ for daily controls or in laboratories for the analysis of series of milk samples. This test does not require any sample processing, cleaning or extraction.

BENEFITS

- X3 MULTIPLE**
 Detection of β -lactams, Tetracyclines and Sulfonamides.
- RAPID**
 Results in 6 minutes.
- SENSITIVE**
 EU MRLS.
- RELIABLE AND ROBUST**
- USER-FRIENDLY**
 Easily performed on site or in lab.
- COST-EFFECTIVE**

HOW TO USE




SENSITIVITY

	β-LACTAMS	LOD (PPB)	SULFONAMIDES	LOD (PPB)	TETRACYCLINES	LOD (PPB)
PENICILLINS	Penicillin-G	2,5-3,5	Sulfadiazine	8-10	Tetracycline (*)	80-100
	Ampicillin	3-4	Sulfapyridine	0,5-1	Oxytetracycline	60-70
	Amoxicillin	3-4	Sulfathiazole	7,5-8,5	Chlortetracycline	50-60
	Oxacillin	13-15	Sulfamethoxazole	320-360	Doxycycline	20-30
	Cloxacillin	7-9	Sulfamethazine	1-2		
	Dicloxacillin	4-6	Sulfamethoxyipyridazine	2-3		
	Nafcillin	50-70	Sulfadimethoxine	10-15		
CEFALOSPORINS	Ceftiofur	8-12	Sulfacetamide	300-600		
	Cefquinome	16-20	Sulfamerazine	2-3		
	Cefazolin	15-19	Sulfamonomethoxine	8-12		
	Cefapirin	4-6	Sulfaquinoxaline	14-18		
	Cefacetrile	16-20	Sulfachloropyridazine	5-10		
	Cefoperazone	3-4	Sulfaguandine	15-25		
	Cefalexin	700-800	Sulfamethizole	220-260		
	Cefalonium	3-5	Sulfasalazine	250-350		

(*) Other Tetracyclines compounds are detected at a lower concentration than Tetracycline: Minocycline, Demeclocycline, Sancycline, Amicycline, Meclocyline.

RESULTS

Visual interpretation of the result obtained by comparing the intensity of each test line with the control (CTRL) line.



T: Tetracyclines - S: Sulfamides - B: Betalactams

An instrumental reading is also possible with the ReadSensor device.

TriSensor

PRODUCTS REFERENCES

KIT035

TRISENSOR MK BTS MRL - 96 PC

APP032

HEATSENSOR DUO

APP052

HEATSENSOR OCTO



APP038

READSENSOR ESE SMALL CASE

APP039

READSENSOR ESE FULL CASE



Liège Science Park
Allée de la Cense-Rouge, 98
B-4102 Ougrée (Liège)
BELGIUM

www.unisensor.be

Phone +32 4 252 66 02
Fax +32 4 252 90 55
info@unisensor.be

Reseller



Waverssesteenweg 1110/7 Chaussée de Wavre
OUDERGEM 1160 AUDERGHEM - BELGIUM
T: +32 (0)2 663 15 45 - F: +32 (0)2 660 20 98
www.intermed.be - info@intermed.be