

DRYSWAB™ MW813

Product Reference	Product Description	Length of Swab Shaft	Diameter of Swab Shaft	Immediate Packaging	Sterility
MW813	Dryswab™, Polyester PurFlock Fibre, Ultrafine Tip, Flexible Shaft, Breakpoint 100mm	153.4mm with 100mm breakpoint	3.56mm	Individual in Peel Pouch	Sterilised by Gamma Irradiation

PRODUCT DESCRIPTION

Product code MW813 consists of a plastic shaft tipped with polyester flock fibre. The swab has a breakpoint at 100mm. The swab is supplied in a paper and polythene peel pouch for storage prior to use.

ROUTINE QC TESTING SPECIFICATION

Every Bulk batch is tested to ensure that *Haemophilus influenzae* NCTC 8143 is recovered without inhibition after a specified duration of time. Dryswab® products fail QC performance testing if after incubation for a specified duration of time, dependent upon product and organism being tested, the results have dropped by more than 2-log compared to the control tube at 0hr.

REGULATORY

All Dryswab™ are CE-marked and conform to: European Medical Devices Directive (93/42/EEC), 1993, Official Journal of European Communities, L169.

QUALITY ASSURANCE

Medical Wire & Equipment is approved to ISO 9001:2015, ISO 13485:2016 for the design and manufacture of sterile medical devices, and the design and manufacture of sterile and non-sterile in-vitro diagnostic devices.

Code	EAN	Name	Colour	Length	Qty	Pack Size	No. in Case	Shelf Life
MW813				153.4mm				

